



Supplemental Figure 2. The probe-based tissue spectroscopy setup for in-situ measurements. Raman spectra were acquired while inserting the needle from epi-/dermis to spinal cord. A 2 mm outer diameter Raman probe is integrated into the 2 mm inner diameter custom short bevel needle and Raman spectra were collected at the tip of the needle.