Appendix 1. Miya Model Costs

The entire model with a full set of replaceable parts (initial start-up costs) sells for \$6,700. For vaginal hysterectomy, we have two options for parts to use. Our premium vaginal hysterectomy set (the high-fidelity cartridges we used for the study) retail for a combined \$440. The premium vagina costs \$270 and the premium uterus costs \$170. After the hysterectomy is performed, the premium vagina can be used for a variety of other procedures including transobturator slings, retropubic slings, anterior colporrhaphy, posterior colporrhaphy, and sacrospinous ligament suspension. Average cost for these procedures is \$45.00. The uterus can be repurposed to create fibroids to practice operative hysteroscopy bringing the cost per procedure to <\$85. So, while the initial cost for a vaginal hysterectomy with the premium vaginal hysterectomy set is \$440, the cost per procedure can be greatly reduced with the parts being used for other procedures. We also offer our training uterus and vagina for a combined cost of \$218. The training uterus sells for \$99 and the training vagina sells for \$119.

The premium vagina and premium uterus support the following procedures:

- Vaginal Hysterectomy with bleeding
- Anterior Colporrhaphy
- Posterior Colporrhaphy
- Transobturator Sling
- Retropubic Sling
- Colpocleisis
- McCall's culdoplasty

The premium vagina contains the following features:

- Two distinct layers to support a full-thickness vaginal wall dissection
- Realistic cutting, suturing, and puncturing haptics

The premium uterus contains the following features:

- Anatomically correct shape with slight antelexion
- Surgical tubing capable of hooking up to an IV system to simulate bleeding
- Mesh embedded in the ligament aspects of the uterus to support proper suture placement
- Endopelvic fascia marker to signify correct anterior cul de sac entry

The training vagina and training uterus support the following procedures:

- Vaginal Hysterectomy without bleeding
- McCall's culdoplasty

The training vagina contains the following features:

- Single layer in the vagina
- McCall's culdoplasty

The training uterus contains the following features:

- Anatomically correct shape with slight antelexion

Chen CCG, Lockrow EG, Destephano C, Nihira MA, Matthews C, Kammire L, et al. Establishing validity for a vaginal hysterectomy simulation model for surgical skills assessment. *Obstet Gynecol* 2020;136.

The authors provided this information as a supplement to their article.

©2020 American College of Obstetricians and Gynecologists.

Appendix 2. Vaginal Surgical Skills Index

Please circle the number corresponding to the candidate's performance in each category, **irrespective of training level**

1. Initial inspection (Check one)	☐ ₀ Incomplete and unsystematic inspection of relevant pelvic and vaginal structures	☐ ₁ Partially complete and unsystematic inspection of relevant pelvic and vaginal structures	☐ ₂ Complete but unsystematic inspection of relevant pelvic and vaginal structures	☐ ₃ Complete and somewhat systematic inspection of relevant pelvic and vaginal structures	☐ ₄ Systematic and complete assessment of relevant pelvic and vaginal structures	☐ Not observed
2. Incision (Check one)	☐ ₀ Does not perform appropriate incision(s) safely and does not use incision(s) effectively ensuring optimal exposure	☐ ₁ Incompletely performs appropriate incision(s) safely and does not use incision(s) effectively ensuring optimal exposure	☐ ₂ Performs appropriate incision(s) safely but does not use incision(s) effectively ensuring optimal exposure	☐ ₃ Performs appropriate incision(s) safely and partially uses incision(s) effectively ensuring optimal exposure	☐ ₄ Performs appropriate incision(s) safely and uses incision(s) effectively ensuring optimal exposure	☐ Not observed
3. Maintenance of visibility (Check one)	☐ ₀ Almost never or never obtains appropriate exposure	☐ ₁ A few times (less than half the time) obtains appropriate exposure	☐ ₂ Sometimes (about half the time) obtains appropriate exposure	☐ ₃ Most time (more than half the time) obtains appropriate exposure	☐ ₄ Almost always or always obtains appropriate exposure	☐ Not observed
4. Use of assistants (Check one)	☐ ₀ Almost never or never strategically uses assistant(s) to the best advantage	☐ ₁ A few times (less than half the time) strategically uses assistant(s) to the best advantage	☐ ₂ Sometimes (about half the time) strategically uses assistant(s) to the best advantage	☐ ₃ Most time (more than half the time) strategically uses assistant(s) to the best advantage	☐ ₄ Almost always or always strategically uses assistant(s) to the best advantage	☐ Not observed
5. Knowledge of instruments (Check one)	☐ ₀ Almost never or never uses and is familiar with correct instruments	☐ ₁ A few times (less than half the time) uses and is familiar with correct instruments	☐ ₂ Sometimes (about half the time) uses and is familiar with correct instruments	☐ ₃ Most time (more than half the time) uses and is familiar with correct instruments	☐ ₄ Almost always or always uses and is familiar with correct instruments	☐ Not observed
6. Tissue and instrument	☐ ₀ Almost never or never	☐ ₁ A few times (less than half the	☐ ₂ Sometimes (about half the	☐ ₃ Most time (more than half the	☐ ₄ Almost always or always handles	☐ Not observed

Chen CCG, Lockrow EG, Destephano C, Nihira MA, Matthews C, Kammire L, et al. Establishing validity for a vaginal hysterectomy simulation model for surgical skills assessment. *Obstet Gynecol* 2020;136.

The authors provided this information as a supplement to their article.

©2020 American College of Obstetricians and Gynecologists.

handling (Check one)	appropriately handles tissue and instruments	time) appropriately handles tissue and instruments	time) handles tissue and instruments appropriately	time) handles tissue and instruments appropriately	tissue and instruments appropriately	
7. Electro-surgery (Check one)	☐ ₀ Almost never or never uses electro-surgery safely and efficiently	☐ ₁ A few times (less than half the time) uses electro-surgery safely and efficiently	☐ ₂ Sometimes (about half the time) uses electro-surgery safely and efficiently	☐ ₃ Most time (more than half the time) uses electro-surgery safely and efficiently	☐ ₄ Almost always or always uses electro-surgery safely and efficiently	☐ Not observed
8. Knot tying /ligation (Check one)	☐ ₀ Almost never or never quickly and correctly performs suture ligation and knot tying	☐ ₁ A few times (less than half the time) quickly and correctly performs suture ligation and knot tying	☐ ₂ Sometimes (about half the time) quickly and correctly performs suture ligation and knot tying	☐ ₃ Most time (more than half the time) quickly and correctly performs suture ligation and knot tying	☐ ₄ Almost always or always quickly and correctly performs suture ligation and knot tying	☐ Not observed
9. Hemostasis (Check one)	☐ ₀ Almost never or never exposes bleeders and uses correct technique to obtain hemostasis safely and effectively	☐ ₁ A few times (less than half the time) exposes bleeders and uses correct technique to obtain hemostasis safely and effectively	☐ ₂ Sometimes (about half the time) exposes bleeders and uses correct technique to obtain hemostasis safely and effectively	☐ ₃ Most time (more than half the time) exposes bleeders and uses correct technique to obtain hemostasis safely and effectively	☐ ₄ Almost always or always exposes bleeders and uses correct technique to obtain hemostasis safely and effectively	☐ Not observed
10. Procedure completion (Check one)	☐ ₀ Almost never or never completely removes fluid and debris and thoroughly inspects for bleeding	☐ ₁ A few times (less than half the time) completely removes fluid and debris and thoroughly inspects for bleeding	☐ ₂ Sometimes (about half the time) completely removes fluid and debris and thoroughly inspects for bleeding	☐ ₃ Most time (more than half the time) completely removes fluid and debris and thoroughly inspects for bleeding	☐ ₄ Almost always or always completely removes fluid and debris and thoroughly inspects for bleeding	☐ Not observed
11. Time and motion (Check one)	☐ ₀ Almost never or never efficiently performs movements with no awkward or	☐ ₁ A few times (less than half the time) efficiently performs movements with no awkward or	☐ ₂ Sometimes (about half the time) efficiently performs movements with no awkward or	☐ ₃ Most time (more than half the time) efficiently performs movements with no awkward or	☐ ₄ Almost always or always efficiently performs movements with no awkward or	☐ Not observed

Chen CCG, Lockrow EG, Destephano C, Nihira MA, Matthews C, Kammire L, et al. Establishing validity for a vaginal hysterectomy simulation model for surgical skills assessment. *Obstet Gynecol* 2020;136.

The authors provided this information as a supplement to their article.

©2020 American College of Obstetricians and Gynecologists.

	unnecessary moves	unnecessary moves	unnecessary moves	unnecessary moves	unnecessary moves	
12. Flow of operation and forward planning (Check one)	☐ ₀ Almost never or never demonstrates forward planning allowing for proper flow of the procedure	☐ ₁ A few times (less than half the time) demonstrates forward planning allowing for proper flow of the procedure	☐ ₂ Sometimes (about half the time) demonstrates forward planning allowing for proper flow of the procedure	☐ ₃ Most time (more than half the time) demonstrates forward planning allowing for proper flow of the procedure	☐ ₄ Almost always or always demonstrates forward planning allowing for proper flow of the procedure	☐ Not observed
13. Knowledge of specific procedure (Check one)	☐ ₀ Almost never or never demonstrates familiarity with all aspects of the operation	☐ ₁ A few times (less than half the time) demonstrated familiarity with all aspects of the operation	☐ ₂ Sometimes (about half the time) demonstrates familiarity with all aspects of the operation	☐ ₃ Most time (more than half the time) demonstrates familiarity with all aspects of the operation	☐ ₄ Almost always or always demonstrates familiarity with all aspects of the operation	☐ Not observed

Overall, is the trainee competent to safely perform this task unsupervised: NO YES

Comments:

Reprinted from Chen C, Korn A, Klingele C, et al. Objective assessment of vaginal surgical skills. Am J Obstet Gynecol 2010;203:79.e1-8. Copyright 2010, with permission from Elsevier.

Chen CCG, Lockrow EG, Destephano C, Nihira MA, Matthews C, Kammire L, et al. Establishing validity for a vaginal hysterectomy simulation model for surgical skills assessment. Obstet Gynecol 2020;136.

The authors provided this information as a supplement to their article.

©2020 American College of Obstetricians and Gynecologists.

Appendix 3. Receiver operating characteristic curve identifying a modified Vaginal Surgical Skills Index cutoff score of 27 (*red arrow*) as 80.3% sensitive (95% CI 69.0–89.0) and 96.7% specific (95% CI 88.0–99.5) for differentiating between competent and noncompetent surgeons at performing vaginal hysterectomy on a simulation model. Diagonal segments produced by ties.