Copyright © by The Journal of Bone and Joint Surgery, Incorporated Isaacs et al. Consequences of Oversizing: Nerve-to-Nerve Tube Diameter Mismatch http://dx.doi.org/10.2106/JBJS.M.01420 Page 1 of 1

TABLE E-1 Nerve Histomorphology						
	Axon Count*		Axon Caliber* (μm)		G-Ratio*	
Group	Experimental	Control	Experimental	Control	Experimental	Control
А	3092 ± 952.6	2922 ± 464.3	2.397 ± 0.2552#	5.653 ± 0.6843	0.5457 ± 0.0179	0.5602 ± 0.0308
В	574 ± 209.0†#	2891 ± 303.8	$3.375 \pm 0.4550 \dagger \#$	5.533 ± 0.8101	$0.7198 \pm 0.0095 \dagger \#$	0.6025 ± 0.0656
С	1437 ± 594.6†‡#	2943 ± 373.7	2.687 ± 0.3120†#	5.585 ± 0.6985	$0.6469 \pm 0.0138 \dagger \#$	0.5889 ± 0.0693
D	2180 ± 781.6††§#	3225 ± 340.3	2.807 ± 0.2983†‡#	5.364 ± 0.8546	0.6692 ± 0.0078†‡#	0.5873 ± 0.0661

^{*}The values are given as the mean and standard deviation. $\dagger P < 0.05$ compared with the value in group A. $\dagger P < 0.05$ compared with the value in group B. $\S P < 0.05$ compared with the value in group C. # P < 0.05 compared with the value in the control group.