

TABLE E-1 95% Confidence Intervals for Data in Figures 2 and 3

	Implant Survival Rate		
	One Year	Three Years	Five Years
All causes of revision			
All-polyethylene	98.7% (96.3% to 100.0%)	98.7% (96.3% to 100.0%)	98.7% (96.3% to 100.0%)
Modular-metal-backed	100.0%	100.0%	98.6% (96.0% to 100.0%)
Rotating-platform	100.0%	98.7% (96.2% to 100.0%)	98.7% (96.2% to 100.0%)
Revision due to loosening			
All-polyethylene	98.7% (96.3% to 100.0%)	98.7% (96.3% to 100.0%)	98.7% (96.3% to 100.0%)
Modular-metal-backed	100.0%	100.0%	98.7% (96.0% to 100.0%)
Rotating-platform	100.0%	100.0%	100.0%