

Table E-1: Life table for implant survival with revision for any reason as the end point.

Time (year)	n.risk	n.event	survival rate	std.err	95% CI	failure rate	95% CI
0	91	0	1.00	0.00	1.00-1.00	0	0.00-0.00
1	80	1	0.99	0.01	0.97-1.00	0.01	0.00-0.03
2	66	5	0.92	0.03	0.87-0.98	0.08	0.02-0.13
3	59	4	0.87	0.04	0.79-0.95	0.13	0.05-0.21
4	51	5	0.79	0.05	0.70-0.89	0.21	0.11-0.30
5	48	2	0.76	0.05	0.67-0.87	0.24	0.13-0.33
6	40	4	0.69	0.06	0.59-0.81	0.31	0.19-0.41
7	31	3	0.63	0.06	0.52-0.77	0.37	0.23-0.48
8	27	2	0.59	0.06	0.48-0.73	0.41	0.27-0.52
9	20	5	0.48	0.07	0.37-0.64	0.52	0.36-0.63
10	15	1	0.45	0.07	0.33-0.61	0.55	0.39-0.67
11	10	2	0.38	0.08	0.25-0.56	0.62	0.44-0.75
12	7	1	0.33	0.08	0.21-0.53	0.67	0.47-0.79
13	5	1	0.29	0.08	0.17-0.50	0.71	0.50-0.83
14	5	0	0.29	0.08	0.17-0.50	0.71	0.50-0.83
...	...	...	...	...	...	...	...
23	2	0	0.29	0.08	0.17-0.50	0.71	0.5-0.83