

Appendix Table A1. Comparison of monocular best corrected visual acuity obtained between EyeNetra refraction and manifest refraction (n=152 eyes) and EyeNetra refraction and cycloplegic refraction (n=74 eyes).

Manifest BCVA			
Netra BCVA	<20/20	≥20/20	P-value*
<20/20	27	57	< .0001
≥20/20	2	66	
	<20/25	≥20/25	<.0001
<20/25	7	53	
≥20/25	2	90	
	<20/40	≥20/40	<.0001
<20/40	1	19	
≥20/40	0	132	
Cycloplegic BCVA			
Netra BCVA	<20/20	≥20/20	P-value*
<20/20	8	29	< .0001
≥20/20	0	37	
	<20/25	≥20/25	< .0001
<20/25	2	21	
≥20/25	0	51	
	<20/40	≥20/40	.03
<20/40	0	5	
≥20/40	0	69	

BCVA=Best-Corrected Visual Acuity; *McNemar's test