



Figure E3: In acute optic neuritis (crosses) and NAION (circles), comparison of commercial 2D segmentation (y axis) and 3D segmentation methods for measuring the GCL+IPL thickness shows the ‘failures’ by using the 2D method. If the failures are ignored the correlation for measurements by each method is strong (solid line).