

Supplemental Digital Content 6. INH dipstick assay compared to commercial test

		IsoScreen N=15	
		Positive n (%)	Negative n (%)
INH dipstick			
Positive	n (%)	7 (87.5)	1 (14.3)
Negative	n (%)	1 (12.5)	6 (85.7)
Sensitivity %		87.5%	
Specificity %			85.7%
Agreement %		86.7%	
Kappa (95% CI)		kappa 0.73, p=0.002	