

Table E1. Comparison of Correct Number of Plates with Eye Handbook (EHB) Color Vision Test and Ishihara Color Plates (ICP) in Control, Study, and Color Vision Deficient (CVD) Subjects

	Controls (N = 76)	Study (N = 113)	CVD ^a (N = 12)
Mean Correct Plates on ICP (SD)	10.92 (0.27)	9.99 (2.30)	4.00 (2.59)
Mean Correct Plates on EHB (SD)	10.93 (0.25)	10.24 (2.25)	4.50 (3.21)
Bland-Altman LOA (95% LOA)	-0.01 (-0.61-0.59)	-0.25 (-1.92-1.42)	-0.50 (-4.64-3.64)

SD = Standard Deviation; LOA = Limits of Agreement

^aCVD subjects were included in the study column.